

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Twsp. PD 2400 Bethel Rd Pennsauken, NJ 08109	2. Contact: Sgt. Sulzbach 3. Phone Number: 856-488-0080 x2301	4. Date: 2-8-2011 5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0066	7. Simulator Component Serial Number:	8. Temperature Probe Component Serial Number:

9. Reason for Service:
EC would not recover after second control test.

10. Comments:
Instrument is being sent back to Draeger for further evaluation.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F. Sullivan #5103
 Name & Badge Number (Print)

Tpr I F Sullivan #5103 2-8-11
 Signature Date

Return & Repair Form:

Customer Information:

Company Name: Remzeuken Tussp B: _____ S: _____
 Date Received: 2/11/2011 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-~~7110~~-6510-6810-7510 Serial #: AR(LM) - 0066
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:			
<u>FL</u>			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FL</u>	<u>1</u>	<u>N/A</u>
<u>mp cal 71</u>	<u>cal</u>	<u>1</u>	<u>N/A</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>N/A</u>

Repair Notes:

FL # AR-01-1723
cal w/ cal i OPS

Service Technician: jc

Date: 2/14/2011

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken TWSP PD 2400 Bethal Rd. Pennsauken, NJ 08109	2. Contact: Sgt. Sulzbach 3. Phone Number: 856-488-0080 x2301	4. Date: 3-1-2011 5. County: Camden
--	--	--

6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARUM-0066</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>
--	--	--

9. Reason for Service:
The Alcotest 7110 came back from outside evaluation from Draeger and is returned to service.

10. Comments:
See attached packing slip for details of Draeger's work conducted on the instrument.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p>Tpr. I F. Sullivan #5103 _____ Name & Badge Number (Print)</p>	<p><i>Tpr. I F. Sullivan #5103</i> _____ Signature</p>	<p>3-1-11 _____ Date</p>
--	--	---

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Pennsauken Twp. PD 2400 Bethel Ave. Pennsauken, NJ	2. Contact: Sgt. Sulzbach	4. Date: 5-1-2012
	3. Phone Number: 856-488-0089	5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0066	7. Simulator Component Serial Number:	8. Temperature Probe Component Serial Number:

9. Reason for Service:
The installation of the Eco-Surge Tripp-Lite surge protector.

10. Comments:
The installation of the Eco-Surge Tripp-Lite surge protector was conducted as per Draegers recommendation. Model number TLP808NETG. The instruments phone line is also routed through the surge protector.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Det. I M. Goncalves #6040
 Name & Badge Number (Print)

Det. I M. Goncalves # 6040
 Signature Date