Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.				
Calibration File No.:	02524	Calib. Date:	08/22/2016	Calib. No.:	00031
Certification File No.:	02525	Cert. Date:	08/22/2016	Cert. No.:	00027
Linearity File No.:	02526	Lin. Date:	08/22/2016	Lin. No.:	00026
Solution File No.:	02578	Soln. Date:	01/07/2017	Soln. No.:	00264
Sequential File No.:	02578	File Date:	01/07/2017		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0339
Control Solution %:	0.100%			Expires:	10/28/2018
Solution Control Lot:	16330			Bottle No.:	0203

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:15S		
Control 1 EC	0.102%	00:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	00:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:16S		
Control 2 EC	0.101%	00:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	00:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18S		
Control 3 EC	0.101%	00:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	00:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI First Name: P MI: E

Signature: Pet P #187 Badge No.: 187

Date: 01/07/2017

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.				
Calibration File No.:	02579	Calib. Date:	01/11/2017	Calib. No.:	00032
Certification File No.:	02580	Cert. Date:	01/11/2017	Cert. No.:	00028
Linearity File No.:	02581	Lin. Date:	01/11/2017	Lin. No.:	00027
Solution File No.:	02591	Soln. Date:	02/06/2017	Soln. No.:	00266
Sequential File No.:	02591	File Date:	02/06/2017		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0338
Control Solution %:	0.100%			Expires:	10/28/2018
Solution Control Lot:	16330			Bottle No.:	0222

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:53S		
Control 1 EC	0.100%	07:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:54S		
Control 2 EC	0.098%	07:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 3 EC	0.099%	07:56S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI First Name: P MI: E

Signature: _____ Badge No.: 187

Date: 02/06/2017

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARUM-0051
Location:	PENNSAUKEN TWSP. P.D.				
Calibration File No.:	02579	Calib. Date:	01/11/2017	Calib. No.:	00032
Certification File No.:	02580	Cert. Date:	01/11/2017	Cert. No.:	00028
Linearity File No.:	02581	Lin. Date:	01/11/2017	Lin. No.:	00027
Solution File No.:	02602	Soln. Date:	03/06/2017	Soln. No.:	00267
Sequential File No.:	02602	File Date:	03/06/2017		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0338
Control Solution %:	0.100%			Expires:	10/28/2018
Solution Control Lot:	16330			Bottle No.:	0357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:07S		
Control 1 EC	0.099%	06:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	06:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:08S		
Control 2 EC	0.098%	06:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:10S		
Control 3 EC	0.098%	06:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FABIANI First Name: P MI: E

Signature: _____ #187

Badge No.: 187

Date: 03/06/2017